

# Performances analysis of piezoelectric cantilever based energy harvester devoted to mesoscale intra-body robot

Kanty RABENOROSOA and Micky RAKOTONDRABE

Automatic Control and MicroMechatronic Systems (AS2M) department  
FEMTO-ST Institute  
CNRS UMR-6174, University of Franche-Comté, ENSMM, UTBM  
24, rue Alain Savary Besançon France;  
rkanty@femto-st.fr ; mrakoton@femto-st.fr

## ABSTRACT

Mesoscale robots, including active capsules, are a promising and well suited approach for minimal invasive intra-body intervention. However, within the numerous works, the main limitation in these robots is the embedded energy used for their locomotion and for the tasks they should accomplish. The limited autonomy and the limited power make them finally unusable for real situations such as active capsules inside body during several tens of minutes. In this paper, we propose an approach to power mesoscale robots by using energy harvesting techniques through a piezoelectric cantilever structure embedded on the robot and through an oscillating magnetic excitation. The physical model of the proposed system is carried out and simulation results are yielded and analyzed accordingly to the influencing parameters such as the number of layers in the cantilever and its dimensions. Finally, the feasibility of this solution is proved and perspectives are discussed.

**Keywords:** Capsules, energy harvesting, magnetic excitation, piezoelectric harvester, mesoscale robots.

## 1. INTRODUCTION

Mesoscale robots and active capsules are clearly identified as future trends of medical robots thanks to their dimensions more suitable for minimal invasive intervention. Among the applications of mesorobots which are essentially capsules, we find: imaging of gastrointestinal tract, biopsy,<sup>1,2</sup> The most successful passive capsule in the literature has been used to achieve monitoring or video acquisition inside a small bowel or GI tract.<sup>3</sup> Nowadays, the development of active capsules is continuously increasing. The main limitation of active capsules, however, is the embedded energy used both for locomotion and for task accomplishment<sup>4,5</sup> despite of the multipurpose robotic abilities.<sup>3,6</sup> This energy limitation is particularly due to the miniature volume of the capsule and thus a difficulty to embed accumulators within this latter.

The use of a direct energy transfer from an external source based on magnetic interaction to move the nano, micro or mesorobots has been investigated.<sup>7-9</sup> Their limitation is mainly the reduced achievable functions as navigation or video acquisition. Further, another energy transfer was proposed to power embedded motors by using inductive coils.<sup>4,10</sup> However, this solution suffers from the limited efficiency of the transmitted energy (less than 5%) and from the non-conformity with the SAR = 0.4 W/Kg (Specific Absorption Rate) prescribed by the ICNIRP (International Commission on Non-Ionizing Radiation Protection) and with the working frequency recommended for human bodies applications.<sup>11</sup> Indeed, the frequency range has to be less than 100 KHz in order to avoid significant absorption of energy and temperature increase.

In order to improve the capability of mesorobots, we propose an alternative solution based on energy harvesting system embedded in the mesorobot. Although several materials (piezoelectric, magnetostrictive, and electroactive polymer, etc.) can be employed to convert mechanical energy to electric one, piezoelectric materials are undeniably the most appreciated for such applications thanks to the high retrievable energy density for a given size<sup>12-14</sup> or if the structure is well designed.<sup>15</sup> In addition, the literature is mainly focused on a vibration energy harvesting for Wireless Sensor Network (WSN).<sup>16-18</sup>

In this paper, we propose a system to overcome the intracorporeal mesorobot limitation of autonomy through energy harvesting. Based on an external oscillating magnetic excitation and a piezoelectric structure, the proposed approach can be used to power capsules mesorobots autonomous for locomotion and for intra-body tasks

accomplishments. The paper is organized as follows. The proposed design for harvesting system is first described in section II. Section 3 is devoted to the presentation and modeling of the magnetic excitation. In section IV, the piezoelectric system for the harvesting is presented and modeled while section V is devoted to the simulation results and discussion. Finally, conclusions and perspectives are discussed in section 7.

## 2. PRESENTATION OF THE WHOLE HARVESTING MODULE

The harvesting module is a part of the mesorobot capsule. The harvester module was designed according to mesoscale robot dimensions in the literature. Typical existing capsules allow us to design a harvesting module enclosed in a cylinder with about 10mm diameter and 40mm length. In figure 1-(a) is shown a Solidworks-CAD scheme of the module. The module has the following elements.

- A cantilever based on piezoelectric material is the basis of the harvesting system. When bended, charge appears on the surface of the cantilever (accordingly to the direct piezoelectric principle). This charge can be afterwards transformed into voltage ( $V^+ - V^-$ ) to power the mesorobot capsule thanks to an electric circuit. A spherical magnetic object, called mass proof, is glued at the extremity of the cantilever. Its diameter is imposed to be  $1mm$ ,  $1.5mm$  or  $2mm$  for an ease of fixation with the cantilever.
- A magnetic source is used as excitation to bend the piezoelectric cantilever through the magnetic object. The magnetic field to be used is oscillating with a frequency nearly similar to the first resonant frequency of the piezoelectric cantilever. At this frequency, the deflection of the latter is high and consequently the yielded charge is important. We assume that it is always possible to orient the external excitation system such that the direction of the magnetic field is parallel to the deflection of the cantilever ( $Z$ -axis). This assumption can be satisfied easily since the harvester module orientation can be controlled by embedded actuators.
- the cantilever is clamped and enclosed in the magnetically permeable cylindrical package as seen in figure 1-(b).




Figure 1. (a) Principle of the harvesting system for mesoscale robot in the constant gradient zone and (b) energy harvester module of the mesorobot.

It is worth to remark that we also use magnetic field because of its biocompatibility relative to other excitation principles (thermal, etc). In addition, its ability has been demonstrated in some medical applications:

- a robot manipulator that interacts with a capsule and that is based on a permanent magnet for generating displacement,<sup>3</sup>
- a modified MRI for micro/mesorobot<sup>7</sup> and actuator for needle insertion,<sup>19</sup>
- other Saddle/Maxwell coil arrangement,<sup>20</sup>
- solenoide arrangement for intraocular microrobot.<sup>21</sup>

### 3. EXCITATION SYSTEM MODELING

In this section, we model the magnetic excitation system in order to provide an analytical formulation of the interaction force with the piezoelectric cantilever and the magnet.

#### 3.1 Expression of the Magnetic Force versus the gradient and Mass Proof

We consider that the capsule stays in a zone with constant magnetic gradient. The mass proof is a soft magnetic body ( $V_p$ ) which interacts to the magnetic field  $B$ . It is possible to drive a soft magnetic body in multi-DOF<sup>15,20,21</sup> but this study is limited to one DOF in order to demonstrate the effectiveness of the proposed harvesting system. Thus the force along  $Z$  is derived as:

$$F_z = n_d V_p M_{sz} \frac{\partial B}{\partial z} \quad (1)$$

where  $n_d$  is the duty cycle,  $V_p$  is the volume of the active material and  $M_{sz}$  is the magnetization.

#### 3.2 Magnetic source modeling

In medical application, the robot has to stay in a zone with constant magnetic gradient and with controllable magnetic field. This requirement can be fulfilled easily by adjusting the design. In fact, recent works have shown that Maxwell coils are able to generate a constant gradient in a specific design.<sup>20</sup> Therefore, by using two Maxwell coils, the expression of the generated magnetic field is:

$$B = \mu_0 N R I^2 \left[ \frac{1}{2 \left( \left( z - \frac{\alpha R}{2} \right)^2 + R^2 \right)} - \frac{1}{2 \left( \left( z + \frac{\alpha R}{2} \right)^2 + R^2 \right)} \right] \quad (2)$$

where  $\alpha$  defines the ratio between the axial distance  $D$  of two coils and their radius  $R$ ,  $I$  is the electric current, and  $N$  is the number of turns. Figure 2 shows the variation of magnetic field and gradient for  $\alpha=1.9$  : the maximum gradient is 29.3 mT/m and 5% of variation is obtained between -0.25m to 0.25m. This zone is large enough to contain a human part: abdomen, head, leg, etc. By decreasing the radius of the coil (and also the distance  $D$ ), it possible to get higher values of the gradient as observed in.<sup>22</sup> The drawback is the reduction of the workspace with constant gradient.




Figure 2. Variation of magnetic field and its gradient for  $\alpha = 1.9$ ,  $R=0.5\text{m}$ ,  $\mu_0=1.26\text{e}^{-6}$ ,  $I= 100\text{A}$ , and  $N=5000$ .

#### 3.3 Gradient and force range in medical applications

In the literatures, there were some results about the use of magnetic excitation to move ferromagnetic or soft magnetic bodies in medical applications (see Tab 1). The main issues referred to the maximum value of the gradient, the maximal/minimal volume of the movable soft magnetic body and the force range. In fact, none of these previous works can provide the requirement for the proposed capsule in term of field gradient, force range

(see figure 2) and volume and shape of the mass proof expected in our applications. However, based on the model and analysis in 3.1 and 3.2 and on figure 2, we can derive the interaction force *versus* the gradient as pictured in figure 3, and we can observe from this that it is possible to obtain a force around 1mN by appropriately choosing the gradient and/or the radius of the magnetic body. Consequently, it is possible to design a harvester that will satisfy our requirements.

Table 1. Magnetic excitation parameters for meso/microrobots in medical applications

Reference	Gradient (mT/m)	Dimension (mm)	Magnetization ( $10^6$ A/m)	Force (mN)
23	700	ellipse 0.95 x 0.4	0.5	0.009
7	40	sphere $r=0.75$	1.35	0.85
24	310	ellipse 4.9 x 2.67	0.61	250
21	-	0.8x0.8x1.2	0.5	0.11
19	40	sphere $r=2.5$	1.36	1200




Figure 3. Force  $F_z$  versus the gradient variation with three radius of magnetic body ( $M_{sz}=1.36 \cdot 10^6$  A/m).

#### 4. MODELING THE MULTIMORPH PIEZOELECTRIC CANTILEVER WITH MASS PROOF

In the previous section, we provided the model of the magnetic excitation and we derived an order of value of the magnetic force that acts on the mass proof. In this section, we give the analytical model of the piezoelectric cantilever combined with the mass proof, the objective being to yield the output electricity (in term of charge  $Q$ ) and the deflection of the cantilever versus the excitation force.

##### 4.1 Presentation of the structure

In order to characterize the optimal structure that would provide the best performances in term of energy harvesting, we consider the general case of a piezoelectric cantilever with  $n$  layers. The layers can be of piezoelectric material (called piezolayers) or non-piezoelectric material (called passive layer). A cantilever with  $n$  layers (piezo and passive) where the number of piezolayers is  $n_p$  ( $n_p \leq n$ ) is called  $n_p$ -morph cantilever and has a  $n$ -layered structure.<sup>25,26</sup> For instance, a unimorph with uni-layered cantilever has only one piezolayer (Fig. 4-b); a unimorph bi-layered cantilever has two layers made up of one piezolayer and one passive layer (Fig. 4-a); a bimorph cantilever with 2-layers is made up of two piezolayers (Fig. 4-c); etc. Consider the multimorph piezoelectric cantilever that supports the mass proof as pictured in Fig. 4-a. The active length of the cantilever is denoted  $l$ , the

widths of the different layers are imposed to be constant and equal to  $w$ , the thickness of the  $i^{th}$  layer is denoted  $h_i$  such that the total thickness is  $h = \sum_{i=1}^n h_i$ . The length  $l$  of the cantilever is infinitely high relative to the total thickness  $h$ . It results that the diameter of the mass proof is negligible face to the length of the cantilever as this diameter is in the same order than the thickness. Consequently, we can assume that the application of a force on the object comes to the application of the force at the tip of the cantilever. The application of an external force  $F_z$  at the tip or a moment  $M$  about the out-of-plane  $y$ -axis will result in a deflection  $\delta_z$  of the cantilever and in an apparition of a total charge  $Q$  at the different electrodes of the piezolayers (Fig. 4-b). Remark that the local  $z$ -axis of the cantilever is parallel to the  $Z$ -axis of the magnetic field.




Figure 4. (a): a multimorph piezoelectric cantilever with a mass proof at its extremity and with an external force and moment as excitation. (b): the resulting deflection  $\delta_z$  and resulting total charge  $Q$ .

## 4.2 Governing equations of the multimorph piezoelectric cantilever without mass proof

In this subsection, we derive the equations of the deflection  $\delta_z$  and of the charge  $Q$  when applying a harmonic force  $F_z(t)$  and a harmonic moment  $M(t)$ . We consider that the mass proof is first absent. The model of the piezoelectric cantilever when the stationary regime is reached can be derived from the general form in <sup>25</sup> by studying the deflection at the tip (distance  $l$ ) of the cantilever and by focusing the analysis on the first resonant mode. We obtain:

$$\begin{pmatrix} \delta_z(L, t) \\ Q(L, t) \end{pmatrix} = \begin{pmatrix} Z_{11} & Z_{12} \\ Z_{21} & Z_{22} \end{pmatrix} \begin{pmatrix} F_o \cos(\Omega t - \psi) \\ M_o \cos(\Omega t - \psi) \end{pmatrix} \quad (3)$$

where  $Z = \begin{pmatrix} Z_{11} & Z_{12} \\ Z_{21} & Z_{22} \end{pmatrix}$  is the transmittance (matrix),  $F_o$  (resp.  $M_o$ ) is the amplitude of the harmonic force (resp. moment) and  $\Omega$  is the frequency ( $rad/s$ ). Notice that the force  $F_z$  calculated in section III is used as the amplitude  $F_o$  of the force here. The moment  $M_o$  is such that:  $M_o = F_o \cdot l$ . We have:

$$\begin{cases} Z_{11} = \frac{4}{m} \frac{\Gamma_1 \alpha_F}{\omega^2 \sqrt{(1-\eta^2)^2 + (2\zeta\eta)^2}} \\ Z_{12} = -\frac{4}{m} \frac{\Gamma_1 \alpha_M k}{\omega^2 \sqrt{(1-\eta^2)^2 + (2\zeta\eta)^2}} \\ Z_{21} = -\frac{4}{m} \frac{\Gamma_2 \alpha_F c_p}{\omega^2 \sqrt{(1-\eta^2)^2 + (2\zeta\eta)^2}} \\ Z_{22} = \frac{4}{m} \frac{\Gamma_2 \alpha_M c_p k}{\omega^2 \sqrt{(1-\eta^2)^2 + (2\zeta\eta)^2}} \end{cases} \quad (4)$$

where  $\zeta < 1$  is the attenuation coefficient (damping coefficient) and  $m$  is the mass of the piezoelectric cantilever and coefficient  $\frac{m}{4}$  in (Eq.4) indicates that the model was derived considering a quarter of  $m$  as mass seen at the tip of the cantilever. A precise calculation of mass  $m$  is based on the geometric properties of the different layers

and on their densities  $\rho_i$ :  $m = w l \sum_{i=1}^n \rho_i h_i$ . Coefficient  $\eta$  is the ratio between the excitation frequency  $\Omega$  and the undamped oscillation  $\omega$ :  $\eta = \frac{\Omega}{\omega}$ , such that  $\omega = \frac{(kl)^2}{l^2} \sqrt{\frac{Cl}{m}}$ , where  $kl = 1.8751$  for the considered first mode and where the flexure rigidity  $C$  is defined by:

$$C = \frac{w}{3} \sum_{i=1}^n \frac{1}{s_{11,i}} \left[ 3h_i \left( \bar{z} - \sum_{j=1}^i h_j \right) \left( \bar{z} - \sum_{j=1}^{i-1} h_j \right) + h_i^3 \right] \quad (5)$$

$\bar{z}$  being the distance of the neutral axis from the lower surface of the cantilever and is defined by:

$$\bar{z} = - \frac{\sum_{i=1}^n \frac{h_i^2}{s_{11,i}} - 2 \sum_{i=1}^n \frac{h_i}{s_{11,i}} \sum_{j=1}^i h_j}{2 \sum_{i=1}^n \frac{h_i}{s_{11,i}}} \quad (6)$$

Coefficient  $s_{11,i}$  is the elastic coefficient of the  $i^{th}$  layer. If the layer is a non-piezoelectric material (passive layer), the elastic coefficient corresponds to:  $s_{11,i} = \frac{1}{E_{11}}$ , with  $E_{11}$  being the axial elasticity (Young modulus along  $x$ -axis). We also have:

$$c_p = \frac{w}{2} \sum_{i=1}^n \frac{d_{31,i}}{s_{11,i} h_i} \left[ 2\bar{z} h_i - 2h_i \sum_{j=1}^i h_j + h_i^2 \right] \quad (7)$$

where  $d_{31,i}$  is the transversal piezoelectric coefficient of the  $i^{th}$  piezolayer. If the layer is passive, i.e. non-piezoelectric material, we have:  $d_{31,i} = 0$ .

The remaining coefficients of (Eq.4) are described in Tab 2

Table 2. *Parameters for the transmittance Z.*

Parameter	Name
$\Psi = \arctan \left( \frac{2\zeta\eta}{1-\eta^2} \right)$	phase
$\alpha_F = \frac{\cosh(kl) \sin(kl) - \cos(kl) \sinh(kl)}{\sin(kl) + \sinh(kl)}$	
$\alpha_M = \frac{\sin(kl) \sinh(kl)}{\sin(kl) + \sinh(kl)}$	
$\Gamma_1 = \tilde{c} - \tilde{s} \frac{\tilde{C}}{\tilde{S}}$	the eigenmode
$\Gamma_2 = k \left( \tilde{S} - \tilde{c} \frac{\tilde{C}}{\tilde{S}} \right)$	
$\tilde{S} = \frac{1}{2} (\sinh(kl) + \sin(kl))$	
$\tilde{C} = \frac{1}{2} (\cosh(kl) + \cos(kl))$	
$\tilde{s} = \frac{1}{2} (\sinh(kl) - \sin(kl))$	
$\tilde{c} = \frac{1}{2} (\cosh(kl) - \cos(kl))$	

### 4.3 Governing equations of the cantilever with mass proof

The different elementary transmittances  $Z_{ij}$  in (Eq.4) were derived when there is no additional mass along or on the piezoelectric cantilever<sup>25</sup>. This calculation was based on the quarter of the mass  $m$  of the cantilever as load at its tip, i.e.  $\frac{m}{4}$ . However, in our case, there is a mass proof  $M_p$  at this tip. So, the total mass seen by an excitation at the tip is  $\frac{m}{4} + M_p$ . Consequently, instead of using (Eq.4), the final transmittances are:

$$\begin{cases} Z_{11} = \frac{1}{\left(\frac{m}{4} + M_p\right)} \frac{\Gamma_1 \alpha_F}{\omega^2 \sqrt{(1-\eta^2)^2 + (2\zeta\eta)^2}} \\ Z_{12} = -\frac{1}{\left(\frac{m}{4} + M_p\right)} \frac{\Gamma_1 \alpha_M k}{\omega^2 \sqrt{(1-\eta^2)^2 + (2\zeta\eta)^2}} \\ Z_{21} = -\frac{1}{\left(\frac{m}{4} + M_p\right)} \frac{\Gamma_2 \alpha_F c_p}{\omega^2 \sqrt{(1-\eta^2)^2 + (2\zeta\eta)^2}} \\ Z_{22} = \frac{1}{\left(\frac{m}{4} + M_p\right)} \frac{\Gamma_2 \alpha_M c_p k}{\omega^2 \sqrt{(1-\eta^2)^2 + (2\zeta\eta)^2}} \end{cases} \quad (8)$$

## 5. SIMULATION AND DISCUSSION

We simulate in this section the model in (Eq.3) and in (Eq.8) in order to analysis the influence of the geometrical properties on the charge  $Q$  and on the deflection  $\delta_z$ . The simulation analysis was carried out using Matlab-Simulink software. The different configurations of the cantilever used for the simulation are illustrated in Fig. 4. We have:

- a unimorph 2-layered piezocantilever (piezocantilever),
- a unimorph 1-layered piezocantilever,
- a bimorph piezocantilever,
- and a  $n$ -morph ( $n = 3 \dots 10$ ) piezocantilever.

The piezoelectric material proposed for this application is the monocrystal PMN-PT (lead magnesium niobate - lead titanate). Its principal advantages are the high coupling factor relative to piezoelectric materials commonly used (such as PZT). Another advantage of the PMN-PT is its compatibility with microfabrication techniques (DRIE, laser, cutting, etc.) in particular when combined with Silicone material.<sup>27</sup> This latter advantage is essential for the packaging aspect because it is possible to fabricate the systems as small as the microfabrication technique is able to do. Though we use PMN-PT in the simulation, other materials (PZT,...) can be used for the model and for the simulation. The numerical values of the used parameters are summarized in Tab. 3.




Figure 5. Different possible piezocantilevers (without mass proof).

Table 3. Numerical values for the simulations.

Cantilever dimensions:	
$l$	15mm, 25mm, 35mm
$w$	2mm, 4mm, 6mm
$h_{pzt}$	see values in the Figures
$h_{si}$	0.2mm for configuration unimorph 2-layers
Excitation:	
$F_0$	1mN
$M_0$	$F_0 \cdot l$
Mass proof ( $M_p$ )	1mg
Properties of the piezoelectric material PMN-PT:	
$\rho_{piezo}$	8200 Kg/m <sup>3</sup>
$s_{11}$	60e-12 TPa <sup>-1</sup>
$d_{31}$	-1200e-12 pm/V
$\zeta$	0.01
Properties of the silicone:	
$\rho_{Si}$	2330 Kg/m <sup>3</sup>
$s_{11Si}$	1/107e9 m <sup>2</sup> /N
$d_{31Si}$	0

### 5.1 Effect of the number of layers

The influence of the number of layers on the harvester performances is first investigated. In this investigation, the thicknesses  $h_i$  ( $i = 1 \dots n$ ) of the piezolayers are imposed to be equal such that the total thickness  $h = \sum_{i=1}^n h_i$  is increasing with the number  $n$  of layers. The results are shown in Fig. 6 with different possible values of thickness  $h_i$  tested ( $h_i = 0.2mm$ ,  $h_i = 0.3mm$  and  $h_i = 0.4mm$ ). We can observe that the unimorph with one layer provides very negligible charge. In fact, this is due to the absence of (or to the weak) strain along the thickness (expansion/contraction) of the piezolayer, even if the whole cantilever performs high deflection (Fig. 6-b). This weak strain along the thickness comes from the fact that the surfaces (upper and lower) of each layer are weakly constrained. We also observe that a lower value of thickness is favorable to obtain a higher charge. Finally, it is deduced from the figure that the furnished charge decreases with the number of layers; and the most interesting configuration in this investigation is the bimorph structure (2-layers).

In the second investigation, we impose the total thickness  $h$  to be constant whatever the number  $n$  of layers is. The results are pictured in Fig. 7. On the one hand, the deflection is independant from the number of layers except for the unimorph 1-layer case. This is predictable since the stiffness is unchanged when the total thickness  $h$  is left constant when the all layers that compose the cantilever are the same (PMN-PT only). On the other hand, once again we find that the charge is increased when the thickness  $h$  (and consequently the elementary thickness  $h_i$ ) is decreased. It is tempting to choose layers and cantilever with weak thickness. Meanwhile, they are fragile and very difficult to fabricate. In addition, the cantilever has to be rigid enough to support the mass proof.

### 5.2 Effect of the cantilever length and width

Investigation on the cantilever length and width effect is now performed. The results are pictured in Fig. 8. We can note that by increasing the width  $w$  of the length  $l$ , we increase the obtained charge. This is due to the fact that by increasing these two parameters will increase the electrodes surfaces  $l \cdot w$ , and therefore will increase the charge itself. The main limitation to having these parameters high lies on the restricted available volume of the capsule.






Figure 6. (a): evaluation of charge  $Q[C]$  versus the number of layers  $n$  for a given thickness  $h_i$  of the piezolayer(s). (b): evaluation of the deflection  $\delta_z [\mu m]$  versus the number of layers  $n$  for a given thickness  $h_i$  of the piezolayer(s).




Figure 7. . (a): evaluation of charge  $Q[C]$  versus the number of layers  $n$  for a given value of the total thickness  $h$  of the cantilever. (b): evaluation of the deflection  $\delta_z [\mu m]$  versus the number of layers  $n$  for a given value of the total thickness  $h$  of the cantilever.




Figure 8. (a): evaluation of charge  $Q[C]$  versus the number of layers  $n$  for a given length  $l$ . (b): evaluation of charge  $Q[C]$  versus the number of layers  $n$  for a given width  $w$ .

### 5.3 Variation of the resonance frequency

The resonance frequency variation is also studied according to the number of layer and for a fixed thickness  $h_i = 0.2mm$  of piezolayers. We can observe in Fig. 9 that the frequency increases with the number of layer. Furthermore, the longest cantilever  $L = 35mm$  provides the lowest resonance frequency. The study of the

resonance frequency is essential because it is directly linked to the ability of the excitation system. In fact, high frequency needs special design of the electronic part due to the risk of overheating.




Figure 9. Evaluation of the frequency  $\omega$  [rad/s] versus the number of layers  $n$  for a given length.

## 6. PROOF OF CONCEPT

A prototype is tested to prove the feasibility of the concept. It is based on Macro Fiber Composite (MCF) from SMART MATERIAL which is a rectangular piezoelectric ceramic rod sandwiched between layers of adhesive, electrodes and polyamide film. The cantilever can be considered as a unimorph with two layers: one piezoelectric layer and one passive layer. Figure 10-a shows the experimental setup to prove the concept. The setup is composed of:

- the piezoelectric cantilever with a mass proof at its tip. Two different sizes have been used for that. They are presented in Figure 10-b and their characteristics are listed in Table 4 (datas from<sup>28</sup>). The dimensions were chosen to be in the order of the simulation ones. Also, an optical sensor is used to measure the displacement (bending) of the cantilever,
- an electromagnet that excites the proof mass. Its alimentation is supplied by a voltage generator and a voltage/current converter,
- an electrical circuit that transforms the charge appearing on the piezoelectric cantilever’s electrodes into voltage,
- and an oscilloscope to display the recuperated voltage.

Table 4. characteristics of the piezoelectric cantilevers.

MCF 1	$L \times w \times h = 28 \times 7 \times 0.3 \text{ mm}$
$d_{31}$	$-2.1 \times 10^{-10} \text{ C/N}$
MCF 2	$L \times w \times h = 28 \times 14 \times 0.3 \text{ mm}$
$d_{31}$	$-2.1 \times 10^{-10} \text{ C/N}$

The magnetic excitation is controlled to be sinusoidal thanks to the generator, the voltage/current converter and electromagnet. In the meantime the resulting displacement at the cantilever tip is observed, see figure 11-**Top**. The voltage generated by the harvester is measured from the output of the harvester module at different frequencies and with the two MCF. Results are shown on figure 11-**Left** for MCF 1 after 45 minutes and




Figure 10. (a) The experimental setup for energy harvesting power mesoscale robot. (b) The used piezoelectric cantilevers.

figure 11-**Right** for MCF 2 after 30 minutes. First, they clearly confirm that the supplied voltage increases with the excitation frequency, as already demonstrated in the literature. Then, they show that, within the utilized dimensions of piezoelectric cantilevers, the supplied voltage is already promising (some mV) even at low frequency (some Hertz). This means that, by choosing a better configuration of the cantilever within the imposed dimensions which are limited by the sizes of the capsule, it is possible to yield higher voltage. For instance, by using a bimorph based on two-layers cantilever, we can maximize the yielded voltage according to the analysis and simulation carried out in the previous sections.

## 7. CONCLUSIONS AND PERSPECTIVES

The *in-vivo* intervention of mesoscale robots like capsules during MIS (minimal invasive surgery) constitutes the future step of the surgical robotics. It will enable painless and improve maneuverability of tools inside an human body. The approach of energy harvesting is proposed in this paper in order to achieve a wireless power supply of mesoscale robot devoted to such objective. The magnetic excitation is chosen thanks to its ability to generate remote and high force, its biocompatibility, and the possibility to use a MRI (magnetic resonance imaging) as an actuator. The harvester module based on piezoelectric material is modeled. Simulation results are presented with thorough analysis according to the number of layer, the thickness of layer, the cantilever length, and the width. They suggested the efficiency of bimorph cantilever compared to the other configurations of piezoelectric cantilever thanks to the following reasons.

## 8. ACKNOWLEDGMENTS

This work is supported by BQR project from University of Franche Comté. It is also supported by the Labex ACTION project (contract ANR-11-LABX-0001-01) and the French RENATECH network and its FEMTO-ST technological facility.

## REFERENCES

- [1] Carta, R., Thon, J., and Puers, R., "A wireless power supply system for robotic capsular endoscopes," *Sensors and Actuators A: Physical* **162**, 177–183 (2010).




Figure 11. Experimental results of the energy harvesting power mesoscale robot: **Top**: applied voltage for the excitation and the resulting displacement of the cantilever, **Left**: output voltage from the harvester with MCF 1 after 45 minutes, and **Right**: output voltage from the harvester with MCF 2 after 30 minutes.

- [2] Quirini, M., Webster, R. J., Menciassi, A., and Dario, P., “Design and fabrication of a motor legged capsule for the active exploration of the gastrointestinal tract,” in [*IEEE International Conference on Robotics and Automation*], (2007).
- [3] Moglia, A., Menciassi, A., Schurr, M. O., and Dario, P., “Wireless capsule endoscopy: from diagnostic devices to multipurpose robotic systems,” *Biomed Microdevices* **9**, 235243 (2007).
- [4] Lenaerts, B. and Puers, R., “An inductive power link for a wireless endoscope,” *Biosensors and Bioelectronics* **22**, 1390–1395 (2007).
- [5] Dario, P., “Future trends in surgical robotics (part 2),” in [*Summer School for Surgical Robotics*], (2011).
- [6] Toennies, J. L., Tortora, G., Simi, M., Valdastrri, P., and R. J. Webster, I., “Swallowable medical devices for diagnosis and surgery: the state of the art,” *J. Mechanical Engineering Science* **224**, 1397–1414 (2009).
- [7] Martel, S., Mathieu, J.-B., Felfoul, O., Chanu, A., Aboussouan, E., Tamaz, S., and Pouponneau, P., “Automatic navigation of an untethered device in the artery of a living animal using a conventional clinical magnetic resonance imaging system,” *Applied Physics* **90**, 1–3 (2007).
- [8] Valdastrri, P., Quaglia, C., Susilo, E., Menciassi, A., Dario, P., Ho, C. N., Anhoeck, G., and Schurr, M. O., “Wireless therapeutic endoscopic capsule: in vivo experiment,” *Endoscopy* **40**, 979–982 (2008).
- [9] Park, S., Cha, K., and Park, J., “Development of biomedical microrobot for intravascular therapy,” *International Advanced Robotic Systems* **7**, 091–098 (2010).
- [10] Carta, R., Pateromichelakis, N., Thon, J., Sfakiotakis, M., Tsakiris, D., and Puers, R., “A wireless powering system for a vibratory-actuated endoscopic capsule,” in [*Euroensors XXIV*], (2010).
- [11] ICNIRP, “Icnirp guidelines for limiting exposure to time-varying electric, magnetic and electromagnetic fields (up to 300 ghz),” *Health Physics* **74**, 494–522 (1998).

- [12] Roundy, S. and Wright, P. K., “A piezoelectric vibration based generator for wireless electronics,” *Smart Mater. Struct.* **13**, 1131–1142 (2004).
- [13] Sodano, H. A., Inman, D. J., and Park, G., “Comparison of piezoelectric energy harvesting devices for recharging batteries,” *Journal of Intelligent Material Systems and Structures* **16**, 799–807 (2005).
- [14] A. Bartasyte, M. Rakotondrabe, T. B. and Ballandras, S., “From green piezoelectric materials to designed hybrid piezoelectric energy harvesters,” *International Workshop on "Advanced Materials Challenges for Alternative Energy Solutions* (Incheon Korea, April 2014).
- [15] Rakotondrabe, M., “Towards high autonomy energy harvesters based on piezoelectric mems,” *Nanoenergy, (International Conference on Nanoenergy)* (Perugia Italy, July 2013).
- [16] Beeby, S. P., Tudor, M. J., and White, N. M., “Energy harvesting vibration sources for microsystems applications,” *Meas. Sci. Technol.* **17**, 175–195 (2006).
- [17] Ferrari, M., Ferrari, V., Guizzetti, M., Marioli, D., and Taroni, A., “Piezoelectric multifrequency energy converter for power harvesting in autonomous microsystems,” *Sensors and Actuators A* **142**, 329–335 (2008).
- [18] Zhu, Y., Moheimani, S. O. R., and Yuce, M. R., “A 2-dof mems ultrasonic energy harvester,” *IEEE Sensors Journal* **11**, 155–161 (2011).
- [19] Vartholomeos, P., Qin, L., and Dupont, P. E., “Mri-powered actuators for robotic interventions,” in [*IEEE/RSJ International Conference on Intelligent Robots and Systems*], (2011).
- [20] Han, B. H., Park, S., and Lee, S. Y., “Gradient waveform synthesis for magnetic propulsion using mri gradient coils,” *Physics in Medicine Biology* **53**, 4639–4649 (2008).
- [21] Kummer, M. P., Abbott, J. J., Kratochvil, B. E., Borer, R., Sengul, A., and Nelson, B. J., “Octomag: An electromagnetic system for 5-dof wireless micromanipulation,” *IEEE Transactions on Robotics* **26**, 1006–1017 (2010).
- [22] Mathieu, J.-B. and Martel, S., “Magnetic microparticle steering within the constraints of an mri system: Proof of concept of a novel targeting approach,” *Biomed Microdevices* **9**, 801–808 (2007).
- [23] Yesin, K. B., Vollmers, K., and Nelson, B. J., “Modeling and control of untethered biomicrobots in a fluidic environment using electromagnetic fields,” *International Journal of Robotics Research* **25**, 527–536 (2006).
- [24] Abbott, J. J., Ergeneman, O., Kummer, M. P., Hirt, A. M., and Nelson, B. J., “Modeling magnetic torque and force for controlled manipulation of soft-magnetic bodies,” *IEEE Transactions on Robotics* **23**, 1247–1252 (2007).
- [25] Ballas, R. G., [*Piezoelectric multilayer beam bending actuators: static and dynamic behavior and aspects of sensor integration*], Springer (2007).
- [26] Rakotondrabe, M., “Piezoelectric cantilevered structures: modeling control and measurement/estimation aspects,” *Springer Verlag* .
- [27] Ivan, I. A., Rakotondrabe, M., Agnus, J., Bourquin, R., Chaillet, N., Lutz, P., Ponot, J.-C., Duffait, R., and Bauer, O., “Comparative material study between pzt ceramic and newer crystalline pmn-pt and pzn-pt materials for composite bimorph actuators,” *Review on Advanced Materials Science (RAMS)* **24**, 1–9 (2010).
- [28] company, S.-M., “<http://www.smart-material.com/MFC-product-main.html#>,”